

Certificate of Analysis

has been analyzed for compliance with the air/gas quality portion of the specification:

ISO 9573-1:2010 Compressed Air Contaminants - Particles by Laser Particle Counter (L)

as reported on this certificate for the sample described below.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC Copyright © 2022 Trace Analytics, LLC

Da aria Sandoval

Maria Sandoval Laboratory Director

Results vs ISO 8573-1:2010 Compressed Air Contaminants - Particles by Laser Particle Counter **(L)**

	Limiting Characteristics	Purity Class (B)	Sample Results (3)	Specification Limit	Pass / Fail	Estimate of Uncertainty % (1)	
Particles	Maximum Number of Particles per	$0.1 < d \le 0.5 \mu m$		1766	≤400,000		±10
	Cubic Meter (m3) as a Function of	0.5 < d ≤ 1.0 μm	2(P)(L)(Q)(E)	0	≤6,000	PASS	±10
	Particles Size, d,	$1.0 < d \le 5.0 \mu m$	2(1)(L)(Q)(L)	0	≤100	17100	±10
	in micrometers (µm)	d > 5.0 μm		0	None		±10
	By Mass Concentration (CP), mg/m ³	(F)	< 0.011			±4.3	
Water	Pressure Dew Point, °C (4)	2	<-53	≤-40	PASS	±30	
Oil	Oil Aerosol, mg/m³		< 0.011			±4.6	
	Oil Vapor, mg/m³		< 0.007			±6.3	
	Total Oil, mg/m³	2	< 0.018	≤0.1	PASS	±10.9	
	Other Vapors, mg/m³ (O4), comprise						
	Mixture Detected (01)		0.567			±6.3	
Other (2)							

Sampling Point Identification

Collection Point: Air Comp Room After Dryer + Filters Purification: Molecular Sieve/Desiccant Sampling Schedule: Other

Next Sample Due Approx: Unspecified



Report 21-40077 R1

Contact

To: Ms. Lynette Caulfield Air Handling Equipment Inc. 1389 Riverside Dr. Sidney, OH 45365

Sampled By: Jeremy Garrett

Sampled For:

Sampling Point Identification:

Collection Point: Air Comp Room After Dryer + Filters

Purification: Molecular Sieve/Desiccant

IVESUITS

Purchase Order: TRACE200207

Customer ID:

Sample Date: Thu, Dec 16, 2021

25109

LPC Sampled: Thu, Dec 23, 2021

Received: Wed, Dec 29, 2021 **Analyzed:** Mon, Jan 3, 2022

Reported: Tue, Feb 1, 2022

Customer Comments:

Results Notes

n/a = not applicable n/d = not determined n/p = not provided n/s = not specified None (or 0) indicates <LOQ, Tr = Trace, >LOD & LOQ (1) At the 95% confidence interval as a percent of the specification limit includes sampling and analytical estimates of uncertainty. Measurement uncertainty is not taken into account when reporting Pass/Fail designations. Pass/Fail is based on test results falling within or outside specified limits. (2) Gases named in ISO 8573-6 Table 2 and/or other measurands required by the specification or customer. (3) Results apply to the sample as received from the customer. (4) Detector tube readings performed by the customer.

Specification Notes

(P) The $0.3 \le d \le 0.5$ µm size range particle results were compared to the ISO 8573-1: 2010 limits for the $0.1 \le d \le 0.5$ µm size range as the Setra LPC can only size down to 0.3 µm. (L) The particle results were obtained using a Setra Systems, Inc. laser particle counter (LPC) model SPC 83016; serial number 3034, calibration on 6/1/2021, next calibration due 6/1/2022. Laser Particle Counter size channels do not exactly align with ISO 8573 ranges. LPC size channels are matched to closest report ranges.

- (Q) No lower limit of quantitation has been determined for the laser particle counter
- (E) The estimate of uncertainty for the laser particle counter data is based on the allowable range for counter efficiency for the $0.5~\mu m$ size channel.

Laboratory Notes

(R1) This revision corrects analytical data from that in the original report. The cause for this revision: Pressure dew point results have been replaced with the results of a detector tube test taken on 1/14/2022 at the request of the customer. (F) Particles by mass concentration have been determined by filter analysis. (04) Other compounds that are not 0il Vapor were detected and may be considered Organic Solvents. The first number in parentheses indicates the goodness of fit of the sample mass spectrum to the NIST library, with 100 being complete agreement.

(01) A complex mixture was identified, which was composed of compounds that are not oil vapor. Poor resolution precluded an exhaustive identification of all mixture components. The following components have been identified. Cyclohexanone, 2-methyl-5-(1-methylethenyl)-, trans(53),Cyclohexane, cyclopropyl- (35), Tetradecane, 1-chloro- (35), Undecanal(38), 3-Methyl-2-(2-oxopropyl)furan (39). The mixture was also identified in the sample blank.

Sampling Collection Conditions				Sampling Collection Information							
	Temperati	ure Pi	essure	Parameter	Madia Na	Flow Rate	Time	Blank	For Detector Tubes Only		
	Tomporati			1 31 311 311	Media No.	(L / min)			Reading	Scale	
Sampling Point	n/p	1	08 psig	Aerosol (Particles & Oil Aerosol)	656325	71	31 : 19				
Ambient	65°F	14	4.7 psia	Water Vapor	5/a-P (6728531)	4	12 : 41		30 mg/m^3	200	
Flowmeter Calibration Information		(Pressure Dew Point)(4)	20/a-P (8103061)		:		mg/m³				
			Calibration Due	Sulfur Dioxide, SO ₂ (4)	0.5/a-P (6728491)				ppmv		
Flowmeter Type	Serial No.				0.5/a-P (CH29401)				ppmv		
Filter Flowmeter				Oil Vapor	942570	4	10 : 13	942624			
Tube Flowmeter				CO, CO ₂ , HC							
NATE OF TAXABLE PARTY O											

NOTE: Blank fields indicate that samples were not obtained for the given limiting characteristic and no analytical results are presented.

Analytical Methods

Test Method	Contaminant	Sampling Technique	Analytical Technique	Accredited	Cal. Cert. No.
CAT-A-01	Gases (CO, CO ₂ , HC)	Gas Collection Bottle	Gas Chromatography - MS / FID	Yes	202203-01
CAT-A-03	Particles by Mass	Membrane Filter (0.2 μm)	Gravimetry	Yes	202203-03
CAT-A-03	Oil Aerosol	Membrane Filter (0.2 μm)	Extraction - Gravimetry	Yes	202203-03
CAT-A-04	Particles by Size	Membrane Filter (0.2 μm)	Optical Microscopy	Yes	202203-04
CAT-A-06	Oil Vapor	Charcoal Tube	Gas Chromatography - Mass Spectrometry	Yes	202203-06
CAT-A-07	Pressure Dewpoint, SO ₂ , NO _x	Gas Detector Tube	Chemical Length-of-Stain	Yes	202203-07
CAT-A-10	Particles by Size	Laser Particle Counter	Laser Particle Counter	Yes	See Laboratory Notes

‡ Trace Analytics, LLC certifies that the instrument(s) associated with the specified method were calibrated in accordance with applicable internal QA procedures.